

EU Declaration of Conformity

Unit type	Fluorescent Array Imager
Model	Fluorowatcher
Serial number	14 digits styled XXXXXYYMMZZZZ, where XXXXXX is model code, YY and MM – year and month of production, ZZZZ – unit number.
Manufacturer	SIA Bioscience Media, Latvia, LV-1002, Riga, Daibes 16.
Applicable Directives	EMC Directive 2014/30/EU LVD Directive 2014/35/EU
Applicable Standards	<u>LVS EN 61326-1: 2013</u> Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements. <u>LVS EN 61010-1: 2011</u> Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements. <u>LVS EN 61010-2-010: 2015</u> Particular requirements for laboratory equipment for the heating of materials. <u>LVS EN 61010-2-051: 2015</u> Particular requirements for laboratory equipment for mixing and stirring.

We declare that this product conforms to the requirements of the above Directives.




Signature

Pavels Bankovskis
CEO

01.10.2019

Date



Signature

Vjaceslavs Smelovs
Electronics Engineer

01.10.2019

Date